

## Certificate

Herewith we, OMRON HEALTHCARE Co., Ltd. certify the blood pressure monitors below are substantially equivalent to BP765 (HEM·7311·ZSA)\* with HEM·RML31 cuff which is clinically validated employs the same algorithm as the predicate device for determining blood pressure.

- HEM-7270 (HEM-7270-AP)
- HEM-7270 (HEM-7270-IN)

\*Hakuo Takahashi, Masamichi Yoshika and Toyohiko Yokoi. Validation of two automatic devices for the self-measurement of blood pressure according to the ANSI /AA MI/IS O81060-2:2009 guidelines: the Omron BP765 (HEM-7311-ZSA) and the Omron BP760N (HEM-7320-Z). Vascular Health and Risk Management 2015:11 49–53.

Very Truly Yours,

March, 11, 2019

Kazunobu Itonaga

General Manager

Product Development Department

Product Development Strategy HQ