



## Certificate

Herewith we, OMRON HEALTHCARE Co., Ltd. Certify HEM-7130 (HEM-7130-TH) is substantially equivalent to HEM-7130 (Japanese model)\* with HEM-CR24 cuff which is clinically validated employs the same algorithm as the predicate device for determining blood pressure.

\*Hakuo Takahashi, Masamichi Yoshika and Toyohiko Yokoi. Validation of three automatic devices for the self-measurement of blood pressure according to the European Society of Hypertension International Protocol revision 2010: the Omron HEM-7130, HEM-7320F, and HEM-7500F. *Blood Pressure Monitoring* 2015, 20:92–97.

Very Truly Yours,

March, 11, 2019

A handwritten signature in blue ink, reading "Kazunobu Itonaga", is written over a horizontal line.

Kazunobu Itonaga

General Manager

Product Development Department

Product Development Strategy HQ