



Certificate

Herewith we, OMRON HEALTHCARE Co., Ltd. certify the blood pressure monitors below are substantially equivalent to BP760N (HEM-7320-Z)*¹ / HEM-7320-LA *² which are clinically validated and employ the same algorithm as the predicate device for determining blood pressure.

- HEM-7346T-AP3(JPN710T)
- HEM-7158T-AP3(JPN610T)
- HEM-7159T-AP3(JPN616T)
- HEM-7156-AP
- HEM-7156T-AP
- HEM-7155T-AP

Very Truly Yours,


*¹ Takahashi H, Yoshika M, Yokoi. Validation of two automatic devices for the self-measurement of blood pressure according to the ANSI/AAMI/ISO81060-2:2009 guidelines: the Omron BP765 (HEM-7311-ZSA) and the Omron BP760N (HEM-7320-Z).

Vascular Health and Risk Management 2015 Jan 9;11:49-53

*² Grover-Páez F1, Cardona-Muñoz EG, Cardona-Müller D, Guzmán-Saldívar VH, Rodríguez-De la Cerda M, Jiménez-Cázarez MB, Totsuka-Sutto SE, Alanis-Sánchez GA, Ramos-Becerra CG. Validation of the Omron HEM-7320-LA, upper arm blood pressure monitor with Intelli Wrap Technology Cuff HEM-FL1 for self-measurement and clinic use according to the European Society of Hypertension International Protocol revision 2010 in the Mexican population.

Blood Pressure monitoring 2017 Dec;22(6):375-378

March, 24, 2021


 Kazunobu Itonaga
 Product Development HQ
 Senior General Manager